



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

HUMANETICS  
23300 Haggerty Road  
Farmington Hills, MI 48335  
Joshua Yingling Phone: 248 778 2153

CALIBRATION

Valid To: September 30, 2024

Certificate Number: 2421.03

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,4</sup>:

I. Mechanical

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Rotary Displacement Transducer	+/- 75 °	0.2 % F.S.	CL-PR-00025C
Linear Displacement Transducers	(0 to 72) mm	0.2 % F.S.	CL-WI-00001C
IR-TRACC	(0 to 120) mm	0.2 % F.S.	CL-PR-000010C CL-PR-00045C
Chest Potentiometer Assemblies	(0 to 90) mm	0.2 % F.S.	SAE J2517: 2016
Force – Load Cells	4 ozf to 400 lbf (50 to 80 000) lbf	0.2 % F.S. 0.2 % F.S.	Class F weights Axial load using load cells
Moment – Load Cells	(50 to 60 000) lbf·in	0.5 % F.S.	Moment load with load cells

Parameter/Range	Frequency	CMC <sup>2,3</sup> (±)	Comments
Acceleration Sensitivity/Frequency Response –  mV/ m/s <sup>2</sup> (1 to 200) m/s <sup>2</sup>  mV/g (0.11 to 20.39) g	(5 to 10) Hz (10 to < 20) Hz (20 to 1000) Hz (> 1000 to 5000) Hz (> 5000 to 10 000) Hz (> 10 000 to 15 000) Hz (> 15 000 to 20 000) Hz	1.0 % of reading 0.70 % of reading 0.50 % of reading 0.70 % of reading 1.5 % of reading 2.5 % of reading 3.0 % of reading	Comparison system using shaker and accelerometers

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> The "Rounded % Value" is chosen to be conservative and larger than the determined "CMC". In the statement of CMC, F.S. represents "Full Scale".

<sup>4</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

## HUMANETICS

*Farmington Hills, MI*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 16<sup>th</sup> day of September 2022.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2421.03  
Valid to September 30, 2024  
Revised October 3, 2022

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*