

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-18446-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 02.02.2022

Date of issue 02.02.2022

Holder of certificate:

mg-sensor GmbH
Airport Boulevard B 210, 77836 Rheinmünster

with the further location:

Knorrstraße 147, EG-351, 80788 München

Calibration in the fields:

Mechanical quantities

- Force
- Torque
- Acceleration

Electrical quantities

DC and low frequency quantities

- DC voltage
- DC current

Thermodynamic quantities

Temperature quantities

- Resistance thermometers
- Direct reading thermometers
- Temperature transmitters,
transducers, data loggers

Humidity quantities

- Devices for relative humidity

Dimensional quantities

Length

- Length measuring instruments

Angle

- Angle of rotation
- Inclination

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories. Laboratories that conform to the requirements of this standard, operate generally in accordance with the principles of DIN EN ISO 9001.

The certificate together with the annex reflects the status as indicated by the date of issue.

The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de/en/accredited-bodies-search.html>.

Abbreviations used: see last page

Page 1 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the accreditation certificate D-K-18446-01-00

Within the measurands/calibration items marked with with * , the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates. The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18446-01-00

Permanent laboratory, Rheinmünster location

| Calibration and Measurement Capabilities (CMC) | | | | | |
|---|---------|-------------|------------------------------------|---|---|
| Measurement quantity / Calibration item | Range | | Measurement conditions / procedure | Expanded uncertainty of measurement ¹⁾ | Remarks |
| Force* Force sensors (safety-belt) | 500 N | to 25 kN | ISO/TS 17242:2014-05 | 1·10 ⁻² | Traction force reference standard measuring device with reference transducer Analogue and digital sensors |
| Force sensors | 2 kN | to 20 kN | DKD-R 3-3:2018 | 2·10 ⁻³ | Compressive force reference standard measuring device with reference transducer Analogue and digital sensors |
| Multi-component force and moment Multi-component transducer (ATD) | 0.05 kN | to < 0.5 kN | KW-F05000:2021 | 2·10 ⁻² | |
| | 0.5 kN | to 25 kN | | 5·10 ⁻³ | |
| | 3 N·m | to < 30 N·m | | 2·10 ⁻² | |
| | 30 N·m | to 1200 N·m | | 5·10 ⁻³ | |
| Force transducer | 0.5 kN | to 600 kN | | 5·10 ⁻³ | |
| Angular velocity Angular velocity sensors | 150°/s | to 3500°/s | KW-AV0002:2014 | 0.5 % | Rotational via incremental encoder for left and right rotation Analogue and digital sensors |
| secondary, dynamic | 8°/s | to 5000°/s | 1 Hz to 200 Hz | 1.5 % / 1.5° | Analogue and digital sensors Calibration result: complex transfer coefficient (analogue: amplitude/phase, digital: amplitude) and indication deviation |

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18446-01-00

Permanent laboratory, Rheinmünster location

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement ¹⁾ | Remarks |
|---|--|---|---|--|
| Acceleration* Acceleration sensors | 200 m/s ² to 20000 m/s ² | Shock excitation DKD-R 3-1 page 2:2018 | 1.2 % | Analogue and digital sensors |
| | 5 m/s ² to 200 m/s ² | Sinusoidal excitation DKD-R 3-1 page 3:2018 10 Hz to 5 kHz > 5 kHz to 10 kHz | 1.2 % / 1.0 ° 2.5 % / 1.5 ° | Analogue and digital sensors Calibration result: complex transfer coefficient (analogue: amplitude/phase, digital: amplitude) and indication deviation |
| Length Displacement sensor (ATD)) | 0 mm to 200 mm | KW-DS0001:05-2021 | 20 µm | Analogue and digital sensors |
| Angle* Angle of rotation Direct rotary encoder systems | 0° to 360° | VDI/VDE 2648 page 1:2009* | 0.2° | Rotation angle sensors |
| | | KW-AN0002:2018 | | Analogue and digital sensors |
| Inclinometers | -90° to 90° | KW-AN0001:2018 | 0.2° | Inclination angle sensors Analogue and digital sensors |
| Temperature* Resistance thermometers, direct reading thermometers, temperature transmitters, transducers and data loggers with resistance sensor (also PTC/NTC) | 10 °C to 50 °C | In the temperature / humidity generator DKD-R 5-1:2018* KW-TE0002:2017 | 0.15 K | Comparative measurement against display of the temperature / humidity generator |
| | | Temperature display devices and simulators, temperature transmitters, transducers and data loggers for base thermocouples (K, N, J) | | DKD-R 5-5:2018* KW-TE0001:2018 |

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-18446-01-00
Permanent laboratory, Rheinmünster location

| Calibration and Measurement Capabilities (CMC) | | | | |
|---|------------------------|---|---|---|
| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement ¹⁾ | Remarks |
| Relative humidity Direct reading electric hygrometers (also data loggers) No psychrometers | 10 % to 80 % | In the temperature / humidity generator Measurement medium air Air temperature: 20 °C to 25 °C KW-HU0001:2017 | 3 % | Comparative measurement against display of the temperature / humidity generator Measurement uncertainty expressed in relative humidity |
| Electrical quantities DC voltage | 0 V 10 mV to 1000 V | KW-VO0001:2017 | 2 µV $1.0 \cdot 10^{-4} U$ | <i>U</i> : measured value Analogue and digital sensors |
| DC current | 0 A to 10 A | KW-CU0001:2017 | $2.0 \cdot 10^{-4} / + 5 \text{ nA}$ | <i>I</i> : measured value |
| DC current current clamps | 0 A to 1000 A | KW-CU0002:2017 1 to <i>N</i> windings | $1.0 \cdot 10^{-2} / + 5 \text{ nA}$ | Analogue and digital sensors |

Permanent laboratory, München location

| Calibration and Measurement Capabilities (CMC) | | | | |
|--|---|---|---|---|
| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement ¹⁾ | Remarks |
| Acceleration* Acceleration sensors, accelerometer measurement chains | 200 m/s ² to 2000 m/s ² | Shock excitation DKD-R 3-1 page 2: 2018 | 1.5 % | Analogue and digital sensors |
| Force* Force sensors | 2 kN to 20 kN | DKD-R 3-3:2018 | $2 \cdot 10^{-3}$ | Analogue and digital sensors |
| Multi-component force and moment Multi-component transducer (ATD) | 0.05 kN to < 0.5 kN | KW-FO5000:2021 | $2 \cdot 10^{-2}$ | Compressive force reference standard measuring device with reference transducer |
| | 0.5 kN to 25 kN | | $5 \cdot 10^{-3}$ | |
| | 3 N·m to < 30 N·m | | $2 \cdot 10^{-2}$ | |
| | 30 N·m to 1200 N·m | | $5 \cdot 10^{-3}$ | Analogue and digital sensors |

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Abbreviations used:

| | |
|-------|--|
| CMC | Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten) |
| DKD-R | Richtlinie des Deutschen Kalibrierdienstes (DKD), herausgegeben von der Physikalisch-Technischen Bundesanstalt |
| KW- | calibration procedure of the mg-sensor GmbH |
| VDE | Verband der Elektrotechnik, Elektronik und Informationstechnik e.V. |
| VDI | Verein Deutscher Ingenieure e.V. |

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.