

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-18446-01-01 according to DIN EN ISO/IEC 17025:2018

Valid from: 22.10.2025

Date of issue: 22.10.2025

This annex is part of the Accreditation Certificate D-K-18446-01-00.

Holder of the Accreditation Certificate:

Humanetics Europe GmbH
Im Breitspiel 6 A, 69126 Heidelberg

with the locations

Humanetics Europe GmbH
Airport Boulevard B210, 77836 Rheinmünster

Humanetics Europe GmbH
Knorrstraße 147, EG-351, 80788 München

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

*This annex to the certificate was issued by the Deutsche Akkreditierungsstelle GmbH (DAkkS) and is digitally sealed.
This annex to the certificate is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any valid and surveyed accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH (www.dakks.de).*

Annex to the Accreditation Certificate D-K-18446-01-01

Calibration in the fields:

Mechanical quantities

- **Force**
- **Acceleration**
- **Pressure**

Thermodynamic quantities**Temperature quantities**

- **Resistance thermometers**
- **Direct reading thermometers**
- **Temperature transmitters, data loggers**

Humidity quantities

- **Devices for relative humidity**

Within the measurands/calibration items marked with *, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates. The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

Valid from: 22.10.2025

Date of issue: 22.10.2025

page 2 of 5

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-18446-01-01
Permanent laboratory, Rheinmünster location
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Force* Force sensors (safety-belt)	500 N to 25 kN	ISO/TS 17242:2014-05	$1 \cdot 10^{-2}$	Traction force reference standard measuring device with reference transducer Analogue and digital sensors
Force sensors	0,5 kN to 25 kN	DKD-R 3-3:2018	$5 \cdot 10^{-3}$	Compressive force reference standard measuring device with reference transducer Analogue and digital sensors
Multi-component force and moment Multi-component transducer (ATD)	0.1 kN to < 0.5 kN	KW-F05000:2024	$1 \cdot 10^{-2}$	
	0.5 kN to 600 kN		$5 \cdot 10^{-3}$	
	2 N-m to 12000 N-m		$5 \cdot 10^{-3}$	
Angular velocity Angular velocity sensors	150°/s to 3500°/s	KW-AV0002:2021	0.5 %	Rotational via incremental encoder for left and right rotation Analogue and digital sensors
secondary, dynamic	8°/s to 5000°/s	KW-AV0005:2021 1 Hz to 200 Hz	1.5 % / 1.5°	Analogue and digital sensors Calibration result: complex transfer coefficient (analogue: amplitude/phase, digital: amplitude) and indication deviation

Valid from: 22.10.2025

Date of issue: 22.10.2025

page 3 of 5

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-18446-01-01
Permanent laboratory, Rheinmünster location
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Acceleration* Acceleration sensors	200 m/s ² to 20000 m/s ²	Shock excitation DKD-R 3-1 page 2:2018	1.2 %	Analogue and digital sensors
	5 m/s ² to 200 m/s ²	Sinusoidal excitation DKD-R 3-1 page 3:2018 10 Hz to 5 kHz > 5 kHz to 10 kHz	1.2 % / 1.0 ° 2.5 % / 1.5 °	Analogue and digital sensors Calibration result: complex transfer coefficient (analogue: amplitude/phase, digital: amplitude) and indication deviation
Pressure* Positive pressure	0 bar to 6 bar	DKD-R 6-1:2014	1 %	
Absolute pressure	1 bar to 7 bar	DKD-R 6-1:2014	1 %	
Temperature* Resistance thermometers, direct reading thermo- meters, temperature transmitters and data loggers with resistance sensor	10 °C to 50 °C	DKD-R 5-1:2018 in temperature / humidity generator	0.15 K	Comparison with display of the temperature / humidity generator
Temperature indicators and simulators for base metal thermocouples	-50 °C to 500 °C	DKD-R 5-5:2018	0.2 K	Characteristic curve according to DIN EN 60584:2014
Relative humidity* Direct reading electric hygrometers, data loggers	10 % to 80 %	DKD-R 5-8:2019 in temperature / humidity generator Measurement in air Air temperature: 20 °C to 25 °C	3 %	Comparison with display of the temperature / humidity generator Measurement uncertainty expressed as absolute value of the relative humidity

Valid from: 22.10.2025

Date of issue: 22.10.2025

page 4 of 5

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-18446-01-01
Permanent laboratory, München location

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Acceleration* Acceleration sensors, accelerometer measurement chains	200 m/s ² to 2000 m/s ²	Shock excitation DKD-R 3-1 page 2: 2018	1.5 %	Analogue and digital sensors
Force* Force sensors	2 kN to 20 kN	DKD-R 3-3:2018	2·10 ⁻³	Analogue and digital sensors
Multi-component force and moment Multi-component transducer (ATD)	0.05 kN to < 0.5 kN	KW-FO5000:2021	2·10 ⁻²	Compressive force reference standard measuring device with reference transducer
	0.5 kN to 25 kN		5·10 ⁻³	
	3 N·m to < 30 N·m		2·10 ⁻²	
	30 N·m to 1200 N·m		5·10 ⁻³	Analogue and digital sensors

Abbreviations used:

DIN	Deutsches Institut für Normung e.V. – German institute for standardization
EN	Europäische Norm – European Standard
IEC	International Electrotechnical Commission
ISO	International Organization for Standardisation
DKD-R	Richtlinie des Deutschen Kalibrierdienstes (DKD), herausgegeben von der Physikalisch-Technischen Bundesanstalt
KW-	calibration procedure of the Humanetics Europe GmbH

Valid from: 22.10.2025

Date of issue: 22.10.2025

page 5 of 5

This document is a translation. The definitive version is the original German annex to the accreditation certificate.