



## *Certificate of Registration*



*This is to certify that Laboratory Management System of*

# **HUMANETICS INNOVATIVE SOLUTIONS INDIA PVT. LTD**

BBMP Khatha No. 937/138 & 139/3B/1, Survey No. 138. Yelahanka, Bellary Road  
Bangalore - 560 064

*Has Been Successfully assessed & comply with the following standard*

### **ISO/IEC 17025:2017**

*Scope of Certification*

*Calibration (For the calibrations covered under this certification, refer to the  
laboratory's Calibration Scope)*

*Initial Issue Date : 31st March 2025\**

*Valid From : 31st March 2026*


*Valid To : 30th March 2027\**

*Certificate No :*

*IAS/IND/12028*

\*Certificate is valid for three years (31st Mar 2025 to 30th Mar 2028) from the date of initial registration. Upon successful completion of the surveillance audit, a new certificate with an extended validity will be issued. For further clarification on the scope of this certification may be obtained by consulting the organization. This certificate is property of IAS and must be returned to IAS, upon cancellation.



  
\_\_\_\_\_  
*Certifying Authority*





## ANNEXURE - SCOPE OF CERTIFICATION ISO/IEC 17025:2017

<b>LABORATORY</b>	HUMANETICS INNOVATIVE SOLUTIONS INDIA PVT. LTD
<b>CERTIFIED STANDARD</b>	ISO/IEC 17025:2017
<b>CERTIFICATE NUMBER</b>	IAS/IND/12028
<b>VALID FROM</b>	MARCH 31, 2026
<b>VALID TO</b>	MARCH 30, 2027*

### CALIBRATION SCOPE – TEST EQUIPMENT SENSORS

Parameter/Equipment	Range	Best Uncertainty <sup>2, 3</sup>	Methods / Standards
Rotary Potentiometers	± 75 degrees	0.2 % F.S.	CL-PR-00025C
Linear Potentiometers	(0 to 72) mm	0.2 % F.S.	CL-WI-00001C
Inclinometers	±37.5 degrees	0.2 % F.S.	CL-PR-00027C
Tilt Sensor	± 45 degrees	0.2 % F.S.	CL-PR-00013C
Knee String Potentiometers	(0 to -38) mm	0.2 % F.S.	CL-PR-00034C

### CALIBRATION SCOPE – LOAD CELLS

Parameter/Equipment	Range	CMC <sup>2, 3</sup> (±)	Comments
Force	4 ozf to 400 lbs 50 to 80000 lbf	0.2 % F.S. 0.2% F.S	Axial load using load cells
Moment	(50 to 60000 in-lbf	0.5 % F.S.	Moment load with load cells

\*Annexure is valid for three years (31<sup>st</sup> March 2025 – 30<sup>th</sup> March 2028) from the date of renewal registration. Upon successful completion of surveillance audit a new annexure with extended validity will be issued. Further clarification regarding the scope of this certification may be obtained by consulting the organization. This certificate is property of IAS, must be returned to IAS, upon certification

\_\_\_\_\_  
Certifying Authority